



HUMAN SERUM QUALITY CONTROL
CERTIFICATE OF ANALYSIS

The Quality Control Department at Randox Laboratories Limited certifies that this product has been tested and meets all performance specifications and release requirements.

Date of Manufacture:	2022-11-29
Product Description:	CALIBRATION SERUM
Catalogue Number:	CAL2351
Batch Number:	1374UE
Expiry Date:	2026-11-28

Analyte	Lot No.	Target	Range		Result	Units
ALBUMIN	1374UE	31.40	N/A	N/A	32.41	g/l
TOTAL PROTEIN	1374UE	48.10	N/A	N/A	47.95	g/l

Human source material from which this product was derived has been tested at donor level for HIV1 & HIV 2, Syphilis, HbsAG, and Antibody to HCV and found to be negative. However, since no method can offer complete assurance as to the absence of infectious agents this material should be handled with the same precautions recommended for a human serum or blood specimen.

Approved By: *Tiernan Sheffington*

Date: *18 Dec 24*

Quality Control Representative