



生化试剂出厂检验报告

Chemistry Reagent Inspection Report

产品名称 Product Name		前白蛋白测定试剂盒（免疫比浊法） Pre-albumin Assay Reagent Kit (Immunoturbidimetric)			
规格 Specification		R1: 45ml×1, R2: 15ml×1, CAL/校准: 1ml×1	产品型号 Product Model	/	
批号 Batch No.		20250403	有效期 Expiration Date	2026/4/2	
生产日期 Production Date		2025. 04. 03	检验日期 Inspection Date	2025. 04. 16	
试剂批号 Reagent batch number		20250403	试剂有效期 Expiration date of Reagent	2026/4/2	
校准/质控批号 Calibrator/Control batch number		20250403	校准有效期 Expiration date of Calibrator	2026/10/2	
检验项目 Inspection Items		检验标准 Inspection Standard		检验结果 Inspection Results	判定 Judgment
1	外观 Appearance	R1无色、R2无色或微黄色，澄清液体，无沉淀及絮状物 R1 is colorless, clear liquid and R2 is colorless or slightly yellowish, semitransparent liquid, without sediments or floccules.		符合 Meet with the requirement	合格 Pass
2	净含量 Net volume	液体试剂的净含量应不少于标示值 The net content of liquid reagents should not be less than the indicated value.		符合 Meet with the requirement	合格 Pass
3	准确度 Accuracy	相对偏差不超过±15% The relative deviation should not exceed ±15%.		3.08%	合格 Pass
4	重复性 Repeatability	变异系数CV≤4% Coefficient of Variation (CV) ≤4%		0.83%	合格 Pass
5	线性范围 Linearity range	在30mg/L~650mg/L (37℃) 范围内，相关系数(r) 应≥0.990 Within the range of 30mg/L~650mg/L (37℃), the correlation coefficient (r) should be ≥0.990.		0.9995	合格 Pass
结论 Conclusion		合格 Pass			
检验员 Inspector		莫宗儒	复核人 Reviewer	高曼姿	
日期 Date		2025. 04. 16	日期 Date	2025. 04. 16	

注：若无内容填写应用“/”。

Note: If there is no information, please fill with "/".

表格编号/Form Number: RT-RFM-Q8503

REV: A. 2/ 保存期限: 10年/Storage Life: 10 years

