



校准物/质控物出厂检验报告

Calibrator/Control Inspection Report

产品名称 Product Name		5'-核苷酸酶测定试剂盒（酶法） 5'-Nucleotidase Assay Reagent Kit (Enzymatic) Control			
规格 Specification		1ml			
批号 Batch No.		20250409	有效期 Expiration Date	20261008	
生产日期 Production Date		2025. 04. 09	检验日期 Inspection Date	2025. 04. 21	
检验项目 Inspection Items		检验标准 Inspection Standard		检验结果 Inspection Results	判定 Judgment
1	外观 Appearance	均匀粉末，无吸潮现象 Uniform powder, nonhygrosc		符合 Meet with the requirement	合格 Pass
2	装量 Net volume	液体试剂的净含量应不少于标示值 The net content of liquid reagents should not be less than the indicated value.		符合 Meet with the requirement	合格 Pass
3	瓶内均匀性 Uniformity	CV(瓶内)≤5% Coefficient of Variation (CV) ≤5%		0.91%	合格 Pass
结论 Conclusion		合格 Pass			
检验员 Inspector		莫宗儒	复核人 Reviewer	高曼安	
日期 Date		2025. 04. 21	日期 Date	2025. 04. 21	

注：若无内容填写应用“/”。

Note: If there is no information, please fill with "/".

表格编号/Form Number: RT-RFM-Q8505

REV: A. 2/ 保存期限: 10年/Storage Life: 10 years