



生化试剂出厂检验报告

Chemistry Reagent Inspection Report

产品名称 Product Name		甘油三酯测定试剂盒 (GPO-PAP法) Triglycerides Assay Reagent Kit (GPO-PAP Method)			
规格 Specification		R1: 50ml×4, R2: 25ml×2	产品型号 Product Model	/	
批号 Batch No.		20250425	有效期 Expiration Date	2026/4/24	
生产日期 Production Date		2025. 04. 25	检验日期 Inspection Date	2025. 04. 30	
试剂批号 Reagent batch number		20250425	试剂有效期 Expiration date of Reagent	2026/4/24	
校准/质控批号 Calibrator/Control batch number		/	校准有效期 Expiration date of Calibrator	/	
检验项目 Inspection Items		检验标准 Inspection Standard		检验结果 Inspection Results	判定 Judgment
1	外观 Appearance	R1无色, R2淡黄色的澄清液体, 无混浊, 无未溶解物 R1 is colorless, clear liquid and R2 is yellowish, clear liquid, without cloudiness or undissolved materials.		符合 Meet with the requirement	合格 Pass
2	净含量 Net volume	液体试剂的净含量应不少于标示值 The Net volume of liquid reagents should not be less than the indicated value		符合 Meet with the requirement	合格 Pass
3	准确度 Accuracy	相对偏差不超过±10% The relative deviation should not exceed ±10%.		2.59%	合格 Pass
4	重复性 Repeatability	变异系数CV≤5% Coefficient of Variation (CV) ≤ 5%.		低值: Low-level: 1.34% 高值: High-level: 0.66%	合格 Pass
5	线性范围 Linearity range	在0.02mmol/L~10mmol/L (37℃) 范围内, 相关系数(r) 应≥0.990 Within the range of 0.02mmol/L ~ 10mmol/L (37℃), the correlation coefficient (r) should be ≥ 0.990.		0.9944	合格 Pass
结论 Conclusion		合格 Pass			
检验员 Inspector		莫宗儒	复核人 Reviewer	高曼姿	
日期 Date		2025. 04. 30	日期 Date	2025/04/30	

注: 若无内容填写应用“/”。

Note: If there is no information, please fill with“/”.

表格编号/Form Number: RT-RFM-Q8503

REV: A. 2/ 保存期限: 10年/Storage Life: 10 years