

化学发光试剂出厂检验报告



Chemiluminescence Reagent Inspection Report

产品名称 Product Name		泌乳素 (PRL) 测定试剂盒 (化学发光免疫分析法) Prolactin (PRL) Reagent Kit (Chemiluminescent Immunoassay)		
规格 Specification		50测试/盒 50Tests/Kit	产品型号 Product Model	/
批号 Batch No.		20250626	有效期 Expiration Date	2026.06.25
生产日期 Production Date		2025.06.26	检验日期 Inspection Date	2025.07.02
检验项目 Inspection Items		检验标准 Inspection Standard		判定 Judgment
1	外观 Appearance	①. 试剂M为棕褐色悬浊液, 久置后, 棕褐色磁珠沉于底部, 上清液为澄明液体; ②. 试剂R为无色澄明液体; ③. 校准品和质控品为无色澄明液体。 ①. The reagent M is a brown brown suspension, after a long time, the brown brown magnetic beads are deposited on the bottom, and the supernatant liquid should be clear liquid; ②. Reagent R is a colorless clear liquid; ③. The calibration and the quality control product for colorless clear liquid.		符合 合格 Meet with the requirement Pass
2	净含量 Net volume	液体试剂的净含量应不少于标示值 The net volume of liquid reagents should not be less than the indicated value.		符合 合格 Meet with the requirement Pass
3	准确度 Accuracy	企业参考品相对偏差不得超过±10%; 质控品相对偏差不得超过±20% The relative deviation of enterprise reference standards should not exceed ±10%; number of quality control product relative deviation less than ±20%.		Q1: -0.36% Q2: 2.79% 准确性样本1 Accuracy1: 5.98% 准确性样本2 Accuracy2: -6.11% 合格 Pass
4	重复性 Repeatability	CV≤8%		低值 Low-level: 3.46% 高值 High-level: 1.11% 合格 Pass
5	线性范围 Linearity range	2.5~4240μ IU/mL内, r≥0.9900 Within the range of 2.5~4240μ IU/mL, the correlation coefficient (r) should be ≥ 0.99.		结果 Result: r= 0.9998 合格 Pass
结论 Conclusion		合格 Pass		
检验员 Inspector		钟宗崇	复核人 Reviewer	高曼姿
日期 Date		2025.07.02	日期 Date	2025.07.02

注: 若无内容填写应用“/”。

Note: If there is no information, please fill with“/”.

表格编号/Form Number: RT-RFM-Q8504

REV: A.2/ 保存期限: 10年/Storage Life: 10 years